

# **FLASH and FIRE POINT APPARATUS (Pensky-Martens Closed Cup) (A473)**

Category: Asphalt

**Product Code:** A473

**Category:** ASPHALT

**Sub-Category:** BITUMEN PROPERTIES

---

- Used to determine the flash point of petroleum products by the Closed Cup Test, with a Flash Point between 40°C to 360°C.
- 

## **STANDARDS**

EN 22719 • ASTM D93 • AASHTO T73

## **ORDERING INFORMATION**

Item	Code
FLASH and FIRE POINT APPARATUS (Pensky-Martens Closed Cup)	A473H00XU

---

*Keywords:*