## **FLASH and FIRE POINT APPARATUS** (Pensky-Martens Closed Cup) (A473)

Category: Asphalt

Product Code: A473

Category: ASPHALT

Sub-Category: BITUMEN PROPERTIES

• Used to determine the flash point of petroleum products by the Closed Cup Test, with a Flash Point between  $40^{\circ}$ C to  $360^{\circ}$ C.

## **STANDARDS**

EN 22719 • ASTM D93 • AASHTO T73

## **ORDERING INFORMATION**

Item Code FLASH and FIRE POINT APPARATUS (Pensky-Martens Closed Cup) A473H00XU

Keywords: